

# Assessment of intraoral scanning technology for multiple implant impressions – A systematic review and meta-analysis

Saloni Kachhara, Deepak Nallaswamy, Dhanraj M. Ganapathy, Vinay Sivaswamy, Vaishnavi Rajaraman

Departments of Prosthodontics and Implant Dentistry, Saveetha Dental College, Saveetha University, SIMATS, Chennai, Tamil Nadu, India

## Abstract

**Background:** Intraoral scanner (IOS) is a medical device used for capturing direct optical impressions and composed of a handheld camera (hardware), a computer and software. Digital impressions by intraoral scanning have become an increasingly popular alternative to conventional impressions. The aim of this systematic review is to assess the studies regarding the various available technologies for IOS and evaluate the most accurate IOS system for cases with multiple implants and identify the factors that can influence its accuracy.

**Materials and Methods:** A comprehensive electronic search was done in online databases, 'Pubmed', 'Google Scholar' and 'Cochrane' based on pre-determined eligibility criteria. In-vitro studies, In-vivo studies and Randomized controlled trials assessing the accuracy of intra-oral scanner technology were selected after thorough screening. The search strategy covered all studies published until February 2019 and yielded a total of 11 articles out of which 8 studies were determined to fulfil the inclusion criteria and were selected for this review. Data extraction from the included studies was conducted by the primary author and reviewed by the second author.

**Results:** The information collected included sample size and population, study design, intervention, scanning methods, comparisons and outcome measures. 5 out of 8 included studies compared the distance deviation of the acquired scans from the true values while the remaining 3 studies gave trueness and precision values as the outcome variables. A forest plot on scanner precision displayed slightly higher precision levels in the TRIOS scanner compared to the other intraoral scanners.

**Conclusion:** Despite the limitations this study, it can be concluded that active wavefront sampling is more accurate than the other intraoral scanning technology employed by commercial scanners.

**Keywords:** Accuracy, active wavefront sampling, confocal microscopy, multiple implant digital impression, optical triangulation, precision, trueness

**Address for correspondence:** Dr. Saloni Kachhara, Department of Prosthodontics and Implant Dentistry, Saveetha Dental College, Saveetha Institute of Medical and Technical Sciences, 162, Poonamallee High Road, Chennai - 600 077, Tamil Nadu, India.  
E-mail: drsbk25@gmail.com

**Submitted:** 15-Oct-2019, **Revised:** 27-Nov-2019, **Accepted:** 08-Feb-2020, **Published:** 07-Apr-2020

## INTRODUCTION

Making an accurate implant impression is a crucial step in fabricating an implant-supported prosthesis.<sup>[1,2]</sup> There

are two methods to make an impression-conventional methods which use elastomeric impression material to record implant position through physical copings and

This is an open access journal, and articles are distributed under the terms of the Creative Commons Attribution-NonCommercial-ShareAlike 4.0 License, which allows others to remix, tweak, and build upon the work non-commercially, as long as appropriate credit is given and the new creations are licensed under the identical terms.

**For reprints contact:** reprints@medknow.com

**How to cite this article:** Kachhara S, Nallaswamy D, Ganapathy DM, Sivaswamy V, Rajaraman V. Assessment of intraoral scanning technology for multiple implant impressions – A systematic review and meta-analysis. J Indian Prosthodont Soc 2020;20:141-52.

Access this article online	
Quick Response Code:	Website: <a href="http://www.j-ips.org">www.j-ips.org</a>
	DOI: 10.4103/jips.jips_379_19

digital implant impressions which use optical methods. Irrespective of the method, the objective is to transfer the intraoral position of dental implants to a working cast or a virtual model.<sup>[2]</sup> Although the conventional impression has been routine in clinical practice for many decades, it is associated with many problems such as material preparation, distortion of impression,

technique sensitivity, time-consuming, and patient discomfort.<sup>[3,4]</sup>

Digital impressions by intraoral scanning have become an increasingly popular alternative to conventional impressions.<sup>[5]</sup> They are a new method for acquiring implant positions and may replace conventional implant impressions and stone cast production.<sup>[6]</sup> Intraoral scanners (IOSs) help in overcoming the mistakes that occur during the conventional impression techniques since no laboratory procedures are involved, and a digital file can be transferred directly into a digital workflow.<sup>[3,7]</sup> Furthermore, IOS impressions help in decreasing the chairside time, enhance patient comfort, and allow for immediately visualizing the adequacy of the impression.<sup>[8,9]</sup>

IOS is a medical device used for capturing direct optical impressions composed of a handheld camera (hardware), a computer and a software.<sup>[10]</sup> The goal of an IOS is to record with precision the three-dimensional (3D) geometry of an object by projecting a light source onto the object to be scanned.<sup>[3,11]</sup> The images captured by imaging sensors are processed by the scanning software, which generates point clouds which are triangulated by the same software, creating a 3D-surface virtual model.<sup>[5]</sup> An increasing number of optical IOSs have been witnessed in the last decade.<sup>[12]</sup> These IOSs are based on different technologies, the choice of which may impact quality of clinical outcome.<sup>[6,11]</sup>




Figure 1: Flowchart of methodology (Prisma chart)

		accordion fringe interferometry) OR interferometry) OR AFI) OR active stereoscopic vision) OR stereoscopy) OR stereoscopy) OR stereoscopy) OR intraoral scanner) OR introral scanner technologies) OR introral scanner technologies) OR intraoral scanners) OR intraoral scanning technology) OR intraoral scanning technologies) OR intraoral scanning technique) OR intraoral scanning techniques) OR triangulation of light) OR optical triangulation) OR continuous imaging) OR video imaging) OR ultrafast optical scanning) OR ultrafast optical sectioning) OR parallel confocal imaging		
#37	Add	Search parallel confocal imaging	385	05:31:07
#36	Add	Search ultrafast optical sectioning	11	05:30:52
#35	Add	Search ultrafast optical scanning	137	05:30:36
#34	Add	Search video imaging	22256	05:30:16
#33	Add	Search continuous imaging	24368	05:30:08
#32	Add	Search optical triangulation	159	05:29:56
#31	Add	Search triangulation of light	146	05:29:24
#30	Add	Search intraoral scanning techniques	353	05:29:05
#29	Add	Search intraoral scanning technique	172	05:28:52
#28	Add	Search intraoral scanning technologies	169	05:28:37
#27	Add	Search intraoral scanning technology	153	05:28:23
#26	Add	Search intraoral scanners	146	05:28:06
#25	Add	Search introral scanner technologies	0	05:27:55
#24	Add	Search intraoral scanner	335	05:27:18
#23	Add	Search stereoscopy	257	05:20:50
#22	Add	Search active stereoscopic vision	369	05:20:39
#21	Add	Search AFI	1021	05:20:16
#20	Add	Search interferometry	22807	05:20:07
#19	Add	Search accordion fringe interferometry	0	05:19:57
#18	Add	Search image processing, computer assisted	222754	05:19:28
#17	Add	Search image processing	145540	05:19:08
#16	Add	Search coherence tomography, optical	39353	05:18:48
#15	Add	Search software	258150	05:18:27
#0	Add	pubmed clipboard	16	07:35:02

Figure 2: PubMed search showing terms for intervention



Kachhara, et al.: Intraoral scanning technology for multiple implant impressions

#80	Add	accuracy, data) OR gap) OR misfit) OR discrepancy) Search (((((((((((((data accuracy) OR reproducibility of results) OR measurement accuracies, dimensional) OR accuracy, dimensional measurement) OR accuracies dimensional measurement) OR precision) OR trueness) OR accuracy) OR data accuracies) OR dimensional measurement accuracies) OR dimensional measurement accuracy) OR accuracies, data) OR accuracy, data) OR gap) OR misfit) OR discrepancy	948734	06:08:13
#79	Add	Search (((((((((((((dental implant) OR dental implants) OR dental implantation) OR dental impression techniques) OR digital implant impressions) OR digital implant impression) OR dental prosthesis, implant-supported) OR multi unit implant impression) OR implant, dental) OR implants, dental) OR dental models) OR scan body) OR scan bodies) OR multiple implant impression	109839	06:06:07
#78	Add	Search (((((((((((((((((((optical coherence tomography) OR confocal microscopy) OR confocal microscopies) OR confocal laser scanning microscopy) OR active wavefront sampling) OR AWS) OR stereophotogrammetry) OR stereophotogrammetries) OR coherence tomography, optical) OR image processing) OR accordion fringe interferometry) OR interferometry) OR AFI) OR active stereoscopic vision) OR stereoscopy) OR intraoral scanner) OR introral scanner technologies) OR optical triangulation) OR triangulation of light) OR intraoral scanning techniques) OR intraoral scanning technique) OR continuous imaging) OR video imaging) OR ultrafast optical scanning) OR parallel confocal imaging)	331791	06:03:23
#77	Add	Search (((((((((((((((((((((((optical coherence tomography) OR microscopy, confocal) OR confocal microscopy) OR confocal microscopies) OR microscopies, confocal) OR confocal laser scanning microscopy) OR laser scanning confocal microscopy) OR microscopy,confocal,laser scning) OR active wavefront sampling) OR AWS) OR stereophotogrammetry) OR stereophotogrammetries) OR tomography, optical coherence) OR OCT tomography) OR software) OR coherence tomography, optical) OR image processing) OR image processing, computer assisted) OR accordion fringe interferometry) OR interferometry) OR AFI) OR active stereoscopic vision) OR stereoscopy) OR stereoscopy) OR intraoral scanner) OR introral scanner technologies) OR introral scanner technologies) OR intraoral scanners) OR intraoral scanning technology) OR intraoral scanning technologies) OR intraoral scanning technique) OR intraoral scanning techniques) OR triangulation of light) OR optical triangulation) OR continuous imaging) OR video imaging) OR ultrafast optical scanning) OR ultrafast optical sectioning) OR parallel confocal imaging)) AND (((((((((((((dental implant) OR dental implants) OR dental implantation) OR maxilla) OR mandible) OR mouth) OR dental impression techniques) OR dental impression technique) OR digital implant impressions) OR digital implant impression) OR dental prosthesis, implant-supported) OR mouth, edentulous) OR multi unit implant impression) OR jaw, edentulous) OR implant, dental) OR implants, dental) OR dental models)) AND (((((((((((((data accuracy) OR reproducibility of results) OR measurement accuracies, dimensional) OR accuracy, dimensional measurement) OR accuracies dimensional measurement) OR speed) OR time) OR trueness) OR precision) OR accuracy) OR trueness) OR data accuracies) OR dimensional measurement accuracies) OR dimensional measurement accuracy) OR accuracies, data) OR gap) OR accuracy, data) OR discrepancy) OR misfit)	8005	05:44:35
#76	Add	Search (((((((((((((data accuracy) OR reproducibility of results) OR measurement accuracies, dimensional) OR accuracy, dimensional measurement) OR accuracies dimensional measurement) OR precision) OR accuracy) OR data accuracies) OR dimensional measurement accuracies) OR dimensional measurement accuracy) OR accuracies, data) OR accuracy, data) OR discrepancy) OR misfit)	4741591	05:44:11

Figure 5: PubMed search showing AND boolein for population, intervention and outcome

Search	Add to builder	Query	Items found	Time
#82	Add	Search (((((((((((((((((((optical coherence tomography) OR confocal mircroscopy) OR confocal microscopies) OR confocal laser scanning microscopy) OR active wavefront sampling) OR AWS) OR stereophotogrammetry) OR stereophotogrammetries) OR coherence tomography, optical) OR image processing) OR accordion fringe interferometry) OR interferometry) OR AFI) OR active stereoscopic vision) OR stereoscopy) OR intraoral scanner) OR introral scanner technologies) OR optical triangulation) OR triangulation of light) OR intraoral scanning techniques) OR intraoral scanning technique) OR continuous imaging) OR video imaging) OR ultrafast optical scanning) OR parallel confocal imaging)) AND (((((((((((((dental implant) OR dental implants) OR dental implantation) OR dental impression techniques) OR digital implant impressions) OR digital implant impression) OR dental prosthesis, implant-supported) OR multi unit Implant impression) OR implant, dental) OR implants, dental) OR dental models) OR scan body) OR scan bodies) OR multiple implant impression)) AND (((((((((((((data accuracy) OR reproducibility of results) OR measurement accuracies, dimensional) OR accuracy, dimensional measurement) OR accuracies dimensional measurement) OR precision) OR trueness) OR accuracy) OR data accuracies) OR dimensional measurement accuracies) OR dimensional measurement accuracy) OR accuracies, data) OR accuracy, data) OR gap) OR misfit) OR discrepancy) Sort by: Best Match Filters: published in the last 5 years	892	07:31:55
#81	Add	Search (((((((((((((((((((optical coherence tomography) OR confocal mircroscopy) OR confocal microscopies) OR confocal laser scanning microscopy) OR active wavefront sampling) OR AWS) OR stereophotogrammetry) OR stereophotogrammetries) OR coherence tomography, optical) OR image processing) OR accordion fringe interferometry) OR interferometry) OR AFI) OR active stereoscopic vision) OR stereoscopy) OR intraoral scanner) OR introral scanner technologies) OR optical triangulation) OR triangulation of light) OR intraoral scanning techniques) OR intraoral scanning technique) OR continuous imaging) OR video imaging) OR ultrafast optical scanning) OR parallel confocal imaging)) AND (((((((((((((dental implant) OR dental implants) OR dental implantation) OR dental impression techniques) OR digital implant impressions) OR digital implant impression) OR dental prosthesis, implant-supported) OR multi unit implant impression) OR implant, dental) OR implants, dental) OR dental models) OR scan body) OR scan bodies) OR multiple implant impression)) AND (((((((((((((data accuracy) OR reproducibility of results) OR measurement accuracies, dimensional) OR accuracy, dimensional measurement) OR accuracies dimensional measurement) OR precision) OR trueness) OR accuracy) OR data accuracies) OR dimensional measurement accuracies) OR dimensional measurement accuracy) OR accuracies, data) OR accuracy, data) OR gap) OR misfit) OR discrepancy)	1527	06:10:43
#80	Add	Search (((((((((((((data accuracy) OR reproducibility of results) OR measurement accuracies, dimensional) OR accuracy, dimensional measurement) OR accuracies dimensional measurement) OR precision) OR trueness) OR accuracy) OR data accuracies) OR dimensional measurement accuracies) OR dimensional measurement accuracy) OR accuracies, data) OR accuracy, data) OR discrepancy) OR misfit)	948734	06:08:13

Figure 6: PubMed search showing final articles after using 5 years filter

to infer distance to the object that is correlated to the focal length of the lens<sup>[13]</sup>

- 3. Active Wavefront Sampling (AWS) – Distance and depth information are derived and calculated from the

pattern produced by each point formed by the rotating module around the optical axis<sup>[14]</sup>

- 4. Stereophotogrammetry – Estimates all coordinates (x, y, and z) only through an algorithmic analysis




Figure 7: Electronic search using Google Scholar

of images, it relies on software and passive light projection.<sup>[15]</sup>

These IOS technologies have their share of clinical impact and pitfalls, which include powdering the surfaces, learning the art of handling the IOS, scanning path to be followed, understanding the tracking and software system.<sup>[10]</sup> Different IOS work on different technologies, and some systems even combine two or more methods to get more accurate scans. The assessment of the accuracy of the impression made by IOS is done by measuring the trueness and precision.<sup>[7]</sup> The purpose of this systematic review is to assess the studies regarding the various available technologies for

IOS and evaluate the most accurate IOS system for scanning multiple implants and identify factors that can influence its accuracy.

### Aim

This systematic review aims to assess various available intraoral scanning methods for multiple implant impressions and evaluate their accuracy.

## MATERIALS AND METHODS

### Structured question

Which is the most explicit intraoral scanning technology for multiple implant impressions in terms of accuracy and precision?

### **PICO (Population, Intervention, Comparison, and Outcomes)**

- P – Multiple implants
- I – Intraoral scanning methods
- C – Nil
- O – Accuracy, trueness, and precision of impression and time taken to make the impression.

### **Outcome variables**

The outcomes of interest in this systematic review are

- Accuracy: Closeness of a measured value to a standard or a known (true) value and to each other (Measured by the difference in distance deviation in  $\mu\text{m}$ )
- Precision: Closeness of measured values between the independent results of the measurement obtained under specific conditions. It measures the repeatability and reproducibility of the results (Measured by difference in distance deviation, implant angulation and depth in  $\mu\text{m}$ )
- Trueness: Trueness is closeness of agreement between the mean obtained from repeated measurements and a true value. It depends on the repeatability of the results (Measured by difference in distance deviation, implant angulation and depth in  $\mu\text{m}$ )
- Speed: Amount of time taken to complete the full mouth scan.

### **Literature search protocol**

Publications of interest within the scope of this focused systematic review were searched in

- The electronic database National Library of Medicine (MEDLINE/PubMed)
- Google Scholar
- Cochrane library
- Web of Science
- EMBASE
- Scopus.

The search was limited to the past 5 years. There were no restrictions or filters applied for the type of literature. A hand search was carried, but no additional articles apart from the electronic search were identified.

### **Search terms**

**P** – Dental implant, dental implants, implants, dental, implant, dental prosthesis, implant-supported, mouth, maxilla, mandible, dental impression technique, humans, dental implant impressions, dental implant impression, dental implantation, jaw, edentulous, multi-unit implant impression, mouth edentulous, and mouth rehabilitation.

**I** – Intraoral scanning technologies, intraoral scanning technology, intraoral scanning technique, intraoral scanning techniques, IOSs, IOS, confocal microscopy, confocal microscopies, confocal laser scanning microscopies, stereophotogrammetry, stereophotogrammetries, optical coherence tomography, software, image processing, video imaging, continuous imaging, ultrafast optical sectioning, ultrafast optical scanning, parallel confocal microscopy, triangulation of light, optical triangulation, accordion fringe interferometry, interferometry, and active stereoscopic vision.

**O** – Accuracy, accuracies, data accuracy, data accuracies, dimensional measurement accuracy, dimensional measurement accuracies, speed, time, trueness, precision, reproducibility of data, repeatability of data, discrepancy, misfit, gap.

### **Article eligibility criteria**

#### *Inclusion criteria*

- Experimental and clinical studies, *in vitro* and *in vivo* studies
- Studies using any one or multiple IOSs
- Articles having outcome measures as accuracy, trueness, or precision
- Studies using digital impressions of multiple implants in edentulous arches.

#### *Exclusion criteria*

- Animal studies
- Studies involving single implant restorations
- Studies involving partially edentulous arches
- Case reports, reviews, systematic reviews
- Studies comparing digital and conventional methods for scanning.

### **Article selection**

#### *Search results*

A total of 1258 articles were obtained using keywords in a Boolean search operator in the PubMed search engine. Duplicates were removed, and the remaining articles were subjected to a title analysis which yielded a total of 16. Further analysis of the articles' abstracts leads to an exclusion of five articles. The remaining 11 articles were subjected to full-text analysis which yielded a total of 8 articles [Figures 1-8].

### **Search strategy**

#### *Data extraction*

The data of the selected studies were extracted using customized data abstraction tables. Information extracted from each study included the following [Tables 1-6]:

- Author and year

- Study design
- Specimen
- Scanning technique
- Scanner
- Implant site and number
- Implant angulation used
- Depth of implant
- Outcome variables
- Sample size
- Scanned surface treatment
- Significance
- Operator.


**RESULTS**

Meta-analysis was planned between two studies namely Stefen *et al.*, 2016 and Mario *et al.*, 2017 as both these studies had a similar methodology, outcome variables and comparison between CEREC Omnicam and Trios 3.

The cumulative results of the meta-analysis display a mild superiority in terms of accuracy for the Trios 3 scanner (AWS) over CEREC Omnicam. Figure 9 shows the meta-analysis of articles comparing the precision of Trios 3 and CEREC Omnicam, and Figure 10 shows a comparison of trueness between the two. Precision of included studies has low heterogeneity (*I*<sup>2</sup>) while the trueness plot is observed to possess high heterogeneity. The overall effect of the consolidated meta-analysis favors the Trios 3 IOS ( $\alpha = 3.53$ ).

**DISCUSSION**

Conventionally, multiple implant impressions are obtained from either direct (open tray/splinted impression) or indirect (closed tray/unsplinted impression) techniques.<sup>[16]</sup> These impressions made from the impression materials have been gold standard for multiple implant impressions for decades.<sup>[17,18]</sup> These impressions are time-consuming,



**Figure 8:** Cochrane search resulted in 2 systematic reviews and 302 trials




Figure 9: Precision




Figure 10: Trueness

messy, and technique sensitive. They lead to many errors and discrepancies in the cast models due to material properties, method of impression making, and laboratory procedures. They are also considered to be uncomfortable for the patients due to various factors such as smell of the material, amount of the material, size of the tray, and the intraoral setting time of the material.<sup>[17]</sup>

The digital revolution has engulfed the dental profession through the introduction of digital impressions through IOSs.<sup>[19,20]</sup> The optical impressions are considered to be relatively more comfortable for the patient and easier and convenient to take for the clinician.<sup>[13,21,22]</sup> They are rapidly overtaking the conventional methods, with the latter likely to disappear in the next few years.<sup>[19]</sup> The last decade has seen an increasing number of optical IOSs, and these are based on different technologies.<sup>[11]</sup>

An IOS is a medical device which takes an optical impression of teeth and implants, using a beam of light.<sup>[23,24]</sup> Irrespective of the type of imaging technology used by IOS, all cameras require the projection of light. This beam of light is then recorded as individual images or video and compiled by the software after recognition of the POI (points of interest). The first two coordinates (x and y) of each point are evaluated on the image, and the third coordinate (z) is then calculated depending on the distance to object technologies of each camera.<sup>[10]</sup> The distance to object technologies is based on principles of optical triangulation, AWS, confocal microscopy, stereophotogrammetry, accordion fringe interferometry or video imaging.

The commercially available scanners based on optical triangulation are CEREC Bluecam; AWS-Lava COS and TrueDef; Confocal microscopy-Trios 3, Trios 3 Mono, iTero, 3D Progress; Video imaging-CS 3600. A few commercial scanners like CEREC Omniscam employ a combination of optical triangulation and confocal microscopy technology.<sup>[5,7]</sup>

The fact that IOS can be a reliable tool for making impressions of single and multiple abutments in patients have been proved by several studies.<sup>[25-27]</sup> However, there is no systematic review compiling the results of these studies based on the technology used in the scanner. This systematic review aims to assess the various technologies used for IOS and the clinical factors affecting it.

In this systematic review, a total of eight *in vitro* studies were evaluated. All the eight studies evaluated the accuracy of the digital impression of the multiple implant casts. They compared the distance deviation in length and angle between the implant scan bodies of the acquired standard tessellation language files from the scanned models to the true values of the master model obtained using an industrial 3D coordinated measurement machine whose accuracy was certified by the National Entity of Accreditation. Five out of eight studies gave the distance deviations from true values and compared the underlying technology by the average error values. The remaining three studies described the trueness and precision of the scanner used.

The average error values obtained for the complete arch multiple implant digital impression from the five included studies were as follows: Lava COS -  $45.02 \pm 37.31 \mu\text{m}$ , Cerec



**Table 1: Characteristics and summary of included studies**

Author and years	Study design	Specimen	Scanning technique	Scanner	Implant site and number	Angulation of implant (°)	Depth of implant (mm)	Scanning method	Sample size	Scanned surface treatment	Operators
Beatriz Gimenez <i>et al.</i> , 2013	<i>In vitro</i>	Edentulous resin model	AWS	Lava COS	12, 22 15, 25 17, 27 6	0, 0 30 distal, 30 mesial 0, 0	4, 2 0, 0 0, 0	Continuous circular scan	50 per group	Application of titanium dioxide powder	4 Group 1-2 experienced Group 2-2 inexperienced
Beatriz Gimenez <i>et al.</i> , 2015	<i>In vitro</i>	Edentulous resin model	Optical triangulation	CEREC bluecam	12, 22 15, 25 17, 27 6	0, 0 30 distal, 30 mesial 0, 0	4, 2 0, 0 0, 0	Continuous scans parallel to the arch	50 per group	-	4 Group 1-2 experienced Group 2-2 inexperienced
Beatriz Gimenez <i>et al.</i> , 2014	<i>In vitro</i>	Edentulous resin model	Parallel confocal laser technology	iTero	12, 22 15, 25 17, 27 6	0, 0 30 distal, 30 mesial 0, 0	4, 2 0, 0 0, 0	BOP	50 per group	Nil	4 Group 1-2 experienced Group 2-2 inexperienced
Beatriz Gimenez <i>et al.</i> , 2015	<i>In vitro</i>	Edentulous resin model	Confocal microscopy	3D progress ZFX intrascan	12, 22 15, 25 17, 27 6	0, 0 30 distal, 30 mesial 0, 0	4, 2 0, 0 0, 0	Continuous around scan bodies	50 per group	Nil	4 Group 1-2 experienced Group 2-2 inexperienced
Beatriz Gimenez <i>et al.</i> , 2015	<i>In vitro</i>	Edentulous resin model	AWS	True definition	12, 22 15, 25 17, 27 6	0, 0 30 distal, 30 mesial 0, 0	4, 2 0, 0 0, 0	-	50 per group	-	4 Group 1-2 experienced Group 2-2 inexperienced
Stefan <i>et al.</i> , 2016	<i>In vitro</i>	Acrylic edentulous mandible model	AWS AWS Active triangulation confocal microscopy	Lava COS True definition CEREC omnicam Trios	36, 46 34, 44 32, 42 6	-	-	-	10 per group	Light powder dusting Light powder dusting Nil Nil	-
Hussam <i>et al.</i> , 2018	<i>In vitro</i>	Edentulous stone model with core structure of tungsten	Confocal microscopy Confocal microscopy Parallel confocal microscopy	Trios 3 Trios 3 mono iTero	12, 22, 23, 25, 15 5	Nonparallel positions	-	BOP BOP Zigzag movement	30 per group	Nil Nil Nil	1 experienced
Mario Imburgia <i>et al.</i> , 2017	<i>In vitro</i>	2 models: Stone model of partially edentulous maxilla Stone model of edentulous maxilla	Active speed 3D video Confocal microscopy and ultrafast Optical scanning Optical triangulation and confocal Microscopy AWS 3D video technology	CS 3600 Trios 3 CEREC omnicam True definition	23, 24, 26 3 11, 21 14, 24 16, 26 6	-	-	-	5 per group	Nil Nil Powder dust	1 experienced

AWS: Active wavefront sampling, 3D: Three-dimensional, BOP: Buccal-occlusal-palatal

Bluecam -  $44.10 \pm 48.5 \mu\text{m}$ , iTero -  $32 \pm 216.1 \mu\text{m}$ , ZFX Intrascan -  $150.6 \pm 1080.3 \mu\text{m}$ , 3DProgress -  $497.4 \pm 1346.0 \mu\text{m}$  and TrueDef -  $26.47 \pm 50.56 \mu\text{m}$ . According to these results, AWS technology gives the least error values followed by confocal microscopy and then optical triangulation. The distance deviation increases with the amount of overlaps taken and also from the first quadrant to the second, with the first scanned quadrant being significantly more accurate than the second.<sup>[28,29]</sup>

The remaining three studies by Stefan *et al.* in 2016, Hussam *et al.*, in 2018, and Imburgia *et al.* in 2017, compared the trueness and precision of the IOSs. Trueness refers to

the closeness of agreement between the expectation of a test result and a true value.<sup>[5,27,30]</sup> Precision is defined as the closeness of agreement between indications or measured quantity values obtained by replicate measurements on the same objects under specified conditions.<sup>[5,27,30]</sup> Ideally, an IOS should have high trueness value, i.e., it should be able to match reality as closely as possible and also high precision value which indicates its repeatability.<sup>[31]</sup> According to Hussam *et al.*, none of the technologies reached the required trueness and precision values and were considered unreliable for multiple implant impression. According to Stefan *et al.*, AWS showed higher trueness and precision compared to confocal microscopy and optical triangulation.

**Table 2: Outcome variables of 5 included studies**

	Beatriz Gimenez <i>et al.</i> , 2013	Beatriz Gimenez <i>et al.</i> , 2015	Beatriz Gimenez <i>et al.</i> , 2014	Beatriz Gimenez <i>et al.</i> , 2015		Beatriz Gimenez <i>et al.</i> , 2015
				ZFX intrscan	3D progress	
Distance deviation (µm)						
Group 1	-29.39±5.49	-28.49±26.91	-14.3±25.6	-32.7±111.1	28.8±94	5.83±12.61
Group 2	-33.35±15.64	-22.46±30.92	-16.2±34.6	-157±292	9.3±29.5	9.86±21.62
Group 3	-45.02±37.31	-107.25±68.65	-27.9±61.6	-142.8±487.7	164.5±526.3	10.05±18.84
Group 4	-11.02±28.12	116.84±94.23	-23.1±148.0	-216.7±836.6	484.6±1057.3	-14.07±33.26
Group 5	-35.28±22.19	-123.09±138.31	-32.0±216.1	-150.6±1080.3	497.4±1346	-26.97±50.56
Implant angulation (µm)						
Angled	-20.2±21.9	-72.7±81.7	-	-125±596	257±776	0.12°±0.05°
Straight	-37.9±26.2	-84.3±99.9	-	-150±693	224±854	0.31°±0.11°
Implant depth (µm)						
Deep	-34.33±18.7	-89.47±105.59	-27.9±61.64	-150±397	87±403	-
Normal	-28.3±29.8	-107.25±68.65	-23.1±149.48	-133±782	337±997	-
Operator experience (µm)						
Experienced	-30.8±25.9	-85.4±98.9	-	-179±601	249±702	-
Inexperienced	13.3±51.2	-47.3±75.7	-	-101±705	224±930	-
Average error	-42.02±37.31	-44.10±48.5	-32±216.1	-150.6±1080.3	497.4±1346.0	-29.97±50.56

3D: Three-dimensional

**Table 3: Outcome variables of 3 included studies**

Author and years	Scanning technology	Scanner	Trueness (µm)	Precision (µm)
Stefan <i>et al.</i> , 2016	AWS	Lava COS	112±25	66±25
	AWS	True definition	35±12	30±11
	Confocal microscopy	Trios	28±7	33±12
	Active triangulation	CEREC omnicam	61±23	59±24
Hussam <i>et al.</i> , 2018	Confocal microscopy	Trios 3	-38	124
	Confocal microscopy	Trios 3 mono	-20	86
	Parallel confocal microscopy	iTero	-35	78
	Active speed 3D video	CS 3600	60.6±11.7	65.5±16.7
Mario Imburgia <i>et al.</i> , 2017	Confocal microscopy and ultrafast optical scanning	Trios 3	67.2±6.9	31.5±9.1
	Optical triangulation and confocal microscopy	CEREC omnicam	66.4±3.9	57.2±9.1
	Active wavefrontsampling 3D video technology	True definition	106.4±23.1	75.3±43.8

AWS: Active wavefront sampling, 3D: Three-dimensional

**Table 4: Enlists the groups of studies based on parameters assessing outcome**

Type of parameter	Total number of studies
Accuracy	5
Precision	3
Trueness	3
Operator experience	6
Implant depth and angulation	7

According to Mario *et al.*, AWS had significantly higher precision and trueness values compared to the others which were almost at similar values. The meta-analysis performed between Trios 3 and CEREC Omnicam for Stefan *et al.* and Mario *et al.* studies favored Trios 3 scanner for better trueness and precision.

There are certain clinical impacts and concerns of digital impressions, especially when it involves full arch implant scanning.<sup>[31]</sup> For digital implant impressions, scan bodies are required which are available separately for every implant size and system which adds to the expense of the impression. The studies included in the review were *in vitro* studies where research was performed on models. Clinically, the oral environment consists of saliva, humidity, limited

mouth opening and also patient anxiety levels adds to the difficulty of impression making.<sup>[4]</sup> Secondary outcome variables such as operator experience, implant angulation and depth, scanning technique were also included in the studies by Beatriz Gimenez *et al.*

Operator experience influences the accuracy of the digital impressions. The accuracy of impressions is better with experienced operator compared to the inexperienced one. However, the inexperienced operators improve the accuracy with the increased number of trials.<sup>[22]</sup> Contrary to this study another study concluded that the performance of the operator is not necessarily dependent on experience.<sup>[32]</sup> However, the author was keen to note that expertise even at the lack of experience, is definitely crucial to the accuracy of digital impression.<sup>[32]</sup>

Implant angulation and depth affect the accuracy of the impression taken. Due to increased angulation of implants, a conventional impression is distorted when the tray is taken out of the mouth,<sup>[33]</sup> and this angulation is limited to 25° for accurate conventional impressions.<sup>[34,35]</sup> Digital impressions made by confocal microscopy technology

**Table 5: Levels of evidence and risk of bias (according to Oxford centre for evidence-based medicine 2011 levels of evidence)**

Author and year	Study design	Level of evidence	Risk of bias
Beatriz Gimenez <i>et al.</i> , 2013	Prospective comparative study	III B	High
Beatriz Gimenez <i>et al.</i> , 2015	Prospective comparative study	III B	High
Beatriz Gimenez <i>et al.</i> , 2014	Prospective comparative study	III B	High
Beatriz Gimenez <i>et al.</i> , 2015	Prospective comparative study	III B	High
Beatriz Gimenez <i>et al.</i> , 2015	Prospective comparative study	III B	High
Wicher <i>et al.</i> , 2012	Prospective comparative study	III B	High
Stefan <i>et al.</i> , 2016	Prospective comparative study	III B	High
Paolo Pesce <i>et al.</i> , 2018	Prospective comparative study	III B	High
Hussam <i>et al.</i> , 2018	Prospective comparative study	III B	High
Leonardo <i>et al.</i> , 2017	Prospective comparative study	III B	High
Mario Imburgia <i>et al.</i> , 2017	Prospective comparative study	III B	High

**Table 6: List of excluded articles**

Author and years	Title	Reason for exclusion
Wicher <i>et al.</i> , 2012	Application of intra-oral dental scanners in the digital workflow of implantology	The outcome measure used was absolute error
Tabea V. Flugge <i>et al.</i> , 2016	Precision of dental implant digitization using intraoral scanners	The cast used was partially edentulous
Paolo <i>et al.</i> , 2018	Precision and accuracy of a digital impression scanner full-arch implant rehabilitation	The outcome measure used was sheffield test

are not significantly affected by the implant angulation or depth of the<sup>[32]</sup> implants.<sup>[36]</sup> The same was reported in another study where the information of the scan bodies in submerged implants was captured sufficiently without affecting the accuracy.<sup>[22]</sup> Angulated implants and the deeply placed implants did not seem to decrease the accuracy in digital impressions.<sup>[22]</sup> The present review observed the implant site and number did not influence the accuracy of impression making using the various intraoral scanning devices. Furthermore, the time and speed of impression making, which is a potential variable that could affect the accuracy of impressions, were not clearly mentioned in the studies included for the systematic review.

The accuracy of full-arch multiple implant scan is related with the correct scanning method. The scanning method and camera movement play an important role in the accuracy of the virtual model.<sup>[37]</sup> Müller *et al.* reported that the zigzag strategy for intraoral scanning has a lower trueness value but a better precision value than buccal–occlusal–palatal strategy.<sup>[38]</sup>

In this systematic review, we could identify only *in vitro* studies. The overall level of evidence is Level 3B; hence, we require well-designed clinical trials with standardized outcomes to recommend the most useful technology and scanner for making an accurate multiple implant digital impression.

## CONCLUSION

Despite the limitations of this study, we can conclude that AWS technique possesses a greater degree of accuracy for making multiple implant digital impression. The degree of expertise of the user is also observed to influence the accuracy of the digital impressions. Implant angulation and depth do not affect the accuracy of digital implants. However, longer clinical trials are required to provide a stronger level of evidence to validate the results of this systematic review.

## Financial support and sponsorship

Nil.

## Conflicts of interest

There are no conflicts of interest.

## REFERENCES

- Wee AG, Aquilino SA, Schneider RL. Strategies to achieve fit in implant prosthodontics: A review of the literature. *Int J Prosthodont* 1999;12:167-78.
- Lee H, Ercoli C, Funkenbusch PD, Feng C. Effect of subgingival depth of implant placement on the dimensional accuracy of the implant impression: An *in vitro* study. *J Prosthet Dent* 2008;99:107-13.
- Ting-Shu S, Jian S. Intraoral digital impression technique: A review. *J Prosthodont* 2015;24:313-21.
- Seo KS, Kim S, Kwon JH, Chang JS. Implant digital impression with intraoral scanners: A literature review implant digital impression with intraoral scanners – A literature review. *Korean Acad Oral Maxillofac Implantol* 2017;21:2-13.
- Imburgia M, Logozzo S, Hauschild U, Veronesi G, Mangano C, Mangano FG. Accuracy of four intraoral scanners in oral implantology: A comparative *in vitro* study. *BMC Oral Health* 2017;17:92.
- Papaspyridakos P, Gallucci GO, Chen CJ, Hanssen S, Naert I, Vandenberghe B. Digital versus conventional implant impressions for edentulous patients: Accuracy outcomes. *Clin Oral Implants Res* 2016;27:465-72.
- van der Meer WJ, Andriessen FS, Wismeijer D, Ren Y. Application of intra-oral dental scanners in the digital workflow of implantology. *PLoS One* 2012;7:e43312.
- Kim J, Park JM, Kim M, Heo SJ, Shin IH, Kim M. Comparison of experience curves between two 3-dimensional intraoral scanners. *J Prosthet Dent* 2016;116:221-30.
- Naidu D, Freer TJ. Validity, reliability, and reproducibility of the iOC intraoral scanner: A comparison of tooth widths and Bolton ratios.

- Am J Orthod Dentofacial Orthop 2013;144:304-10.
10. Richert R, Goujat A, Venet L, Viguie G, Viennot S, Robinson P, *et al.* Intraoral scanner technologies: A review to make a successful impression. *J Healthc Eng* 2017;2017:
  11. Zimmermann M, Mehl A, Mörmann WH, Reich S. Intraoral scanning systems: A current overview. *Int J Comput Dent* 2015;18:101-29.
  12. Rudolph H, Luthardt RG, Walter MH. Computer-aided analysis of the influence of digitizing and surfacing on the accuracy in dental CAD/CAM technology. *Comput Biol Med* 2007;37:579-87.
  13. Taneva E, Kusnoto B, Evans CA. 3D Scanning, imaging, and printing in orthodontics. *Issues Contemp Orthod* 2015;3:148-9.
  14. Logozzo S, Zanetti EM, Franceschini G, Kilpelä A, Mäkynen A. Recent advances in dental optics – Part I: 3D intraoral scanners for restorative dentistry. *Opt Lasers Eng* 2014;54:203-21.
  15. Pradies G, Ferreira A, Özcan M, Giménez B, Martínez-Rus F. Using stereophotogrammetric technology for obtaining intraoral digital impressions of implants. *J Am Dent Assoc* 2014;145:338-44.
  16. Carr AB. A comparison of impression techniques for a five-implant mandibular model. *Implant Dent* 1992;1:232.
  17. Gintaute A, Papatriantafyllou N, Aljehani M, Att W. Accuracy of computerized and conventional impression-making procedures for multiple straight and tilted dental implants. *Int J Esthet Dent* 2018;13:550-65.
  18. Perez-Davidi M, Levit M, Walter O, Eilat Y, Rosenfeld P. Clinical accuracy outcomes of splinted and nonsplinted implant impression methods in dental residency settings. *Quintessence Int* 2016;47:843-52.
  19. van Noort R. The future of dental devices is digital. *Dent Mater* 2012;28:3-12.
  20. Mangano F, Shibli JA, Fortin T. Digital dentistry: New materials and techniques. *Int J Dent* 2016;2016:1-2.
  21. Alghazzawi TF. Advancements in CAD/CAM technology: Options for practical implementation. *J Prosthodont Res* 2016;60:72-84.
  22. Giménez B, Özcan M, Martínez-Rus F, Pradies G. Accuracy of a digital impression system based on active wavefront sampling technology for implants considering operator experience, implant angulation, and depth. *Clin Implant Dent Relat Res* 2015;17 Suppl 1:e54-64.
  23. Yuzbasioglu E, Kurt H, Turunc R, Bilir H. Comparison of digital and conventional impression techniques: Evaluation of patients' perception, treatment comfort, effectiveness and clinical outcomes. *BMC Oral Health* 2014;14:10.
  24. Wismeijer D, Mans R, van Genuchten M, Reijers HA. Patients' preferences when comparing analogue implant impressions using a polyether impression material versus digital impressions (intraoral scan) of dental implants. *Clin Oral Implants Res* 2014;25:1113-8.
  25. Tsirogiannis P, Reissmann DR, Heydecke G. Evaluation of the marginal fit of single-unit, complete-coverage ceramic restorations fabricated after digital and conventional impressions: A systematic review and meta-analysis. *J Prosthet Dent* 2016;116:328-3500.
  26. Ender A, Zimmermann M, Attin T, Mehl A. *In vivo* precision of conventional and digital methods for obtaining quadrant dental impressions. *Clin Oral Investig* 2016;20:1495-504.
  27. Choichlidakis KM, Papaspyridakos P, Geminiani A, Chen CJ, Feng JJ, Ercoli C. Digital versus conventional impressions for fixed prosthodontics: A systematic review and meta-analysis. *J Prosthet Dent* 2016;116:184-90.e12.
  28. Jo SH, Kim KI, Seo JM, Song KY, Park JM, Ahn SG. Effect of impression coping and implant angulation on the accuracy of implant impressions: An *in vitro* study. *J Adv Prosthodont* 2010;2:128.
  29. Giménez B, Özcan M, Martínez-Rus F, Pradies G. Accuracy of a digital impression system based on active triangulation technology with blue light for implants: Effect of clinically relevant parameters. *Implant Dent* 2015;24:498-504.
  30. Ahlholm P, Sipilä K, Vallittu P, Jakonen M, Kotiranta U. Digital versus conventional impressions in fixed prosthodontics: A review. *J Prosthodont* 2018;27:35-41.
  31. Mangano F, Gandolfi A, Luongo G, Logozzo S. Intraoral scanners in dentistry: A review of the current literature. *BMC Oral Health* 2017;17:149.
  32. Giménez B, Özcan M, Martínez-Rus F, Pradies G. Accuracy of a digital impression system based on parallel confocal laser technology for implants with consideration of operator experience and implant angulation and depth. *Int J Oral Maxillofac Implants* 2014;29:853-62.
  33. Jang HK, Kim S, Shim JS, Lee KW, Moon HS. Accuracy of impressions for internal-connection implant prostheses with various divergent angles. *Int J Oral Maxillofac Implants* 2011;26:1011-5.
  34. Assuncao WG, Filho HG, Zaniquelli O. Evaluation of transfer impressions for osseointegrated implants at various angulations. *Implant Dent* 2004;13:358-66.
  35. Assunção WG, Tabata LF, Cardoso A, Rocha EP, Gomes EA. Prosthetic transfer impression accuracy evaluation for osseointegrated implants. *Implant Dent* 2008;17:248-56.
  36. Giménez B, Pradies G, Martínez-Rus F, Özcan M. Accuracy of two digital implant impression systems based on confocal microscopy with variations in customized software and clinical parameters. *Int J Oral Maxillofac Implants* 2015;30:56-64.
  37. Mutwalli H, Braian M, Mahmood D, Larsson C. Trueness and precision of three-dimensional digitizing intraoral devices. *Int J Dent* 2018;2018:1-10.
  38. Müller P, Ender A, Joda T, Katsoulis J. Impact of digital intraoral scan strategies on the impression accuracy using the TRIOS Pod scanner. *Quintessence Int* 2016;47:343-9.